



## **MR<sup>®</sup> ASTM test block**



**Order no.: G57-3 L**

### **Application area**

Test block for Penetrant Testing acc. to ASTM E165 and ASME Boiler and Pressure Vessel Code with declaration of conformity

---

### **Structure**

The ASTM test block is made of aluminum. It is divided into two areas (A and B) by a groove.

---

### **Application**

The test block is intended for testing the display sensitivity of the testing equipment. In area A, the penetrant currently in use is applied, in area B the "fresh" penetrant is applied. By this method a direct comparison is obtained if the batch currently still meets the desired quality and/or fulfills the desired display capability.

**\*\*Technical changes reserved!\*\***